

## **EXPLANATORY NOTES 2 TO THE PROCUREMENT DOCUMENTS**

*pursuant to the stipulations of Section 98 (3), Act 134/2016 Coll., Public Procurement Act, as amended  
(hereinafter referred to as PPA)*

**Name of the public contract:** Photoemission System for Investigation the Surface Electronic Structure of Quantum Materials at the Atomic Level

**Reference number at the Public Procurement Registry:** Z2023-057812

**System ID:** P24V00000280

**Contracting authority:** University of West Bohemia  
with registered seat at: Univerzitní 8, 301 00 Pilsen  
Company reg. no.: 497 77 513

On 21 June 2024, the contracting authority received a request for explanation to the procurement documents containing the following questions to the technical specifications:

### **Question Technical Specifications:**

On page 3 of Annex No. 1, it is stated: "Should allow automated data processing. Adaptation of lens settings when using a lower extraction field, in the range of a few V, as well as 'normal' lens settings of working conditions, when using kV extraction fields."

Do you mean continuous adaptation of the lens settings?

### **Contracting authority's response:**

YES, we mean the continuous adjustment of the lens across the entire range of the extraction voltage setting. This adjustment is a necessary condition for the operando measurement of quantum materials, as defined in the research area (workpackage) "RA3 - Operando characterization and spectroscopy" outlined in the Feasibility Study for the QM4ST project.

The nature of Explanatory notes to the procurement documents shall not constitute an extension for the time limit for submission of tenders.

The time limit for submission of tenders remains unaffected and expires on **9 July 2024 at 10:00 hours**.

Given in Pilsen, see the date in the electronic signature

.....  
**University of West Bohemia**  
Mgr. Kateřina Sladká, MBA  
Legal Department